

## **ABSTRACT**

A method for determining the optical retardation and birefringence values of an anisotropic material utilizing a Fourier transform near infrared spectrophotometer operated in at least a portion of the range of wavenumbers between about 4,000 to about 10,000  $\text{cm}^{-1}$ .

5